GUARDIAN SYSTEMS, INC.

1108 Ashville Road
P.O. Box 190
Leeds, Alabama 35094

Telephone (205) 699-6647
Toll Free (866) 729-7211
Fax (205) 699-3882

City of Leeds
City of Leeds
1040 Park Drive
Leeds, AL 35094
Attention: Mr. Brad Watson

Report Date: 01/18/2019
Receive Date: 01/10/2019
Receive Time: 15:10

Control No : 1901-00156 Sample # 001
Sampler : DB
Sample ID: DC-33-3

Sample Date: 01/10/2019
Sample Time: 9:47

Laboratory Certificate

PARAMETER	RESULTS	UNITS	ANALYST	DATE	TIME	METHOD	REF
pH	7.11	SU	DB	01/10/2019	9:47	150.1	(1)
Turbidity, Nephelometric	10.18	NTU	DB	01/10/2019	9:47	180.1	
Conductance, Sp. (field)	215	umhos	DB	01/10/2019	9:47	120.1	(1)
Nitrogen, Ammonia	<0.1	mg/L	TLB	01/16/2019	15:41	350.2	(1)
Color, APHA	70.	PCU	AH	01/11/2019	14:30	110.2	(1)
Hardness, Total	91.	mg/L	DRH	01/14/2019	12:00	200.7	
Calcium	28.8	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Magnesium	4.55	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Potassium	1.47	mg/L	DRH	01/14/2019	12:00	200.7	(1)

This Certificate is Continued On Next Page.

METHOD REFERENCES

1. Methods for Chemical Analysis of Water and Wastes. EPA-600/4-79-20, revised March 1983, August 1993 May 1994
2. Standard Methods for the Examination of Water and Waste Water, 18th, 19th, 20th, and 22nd Edition, 2012
3. Test Methods for Evaluating Solid Wastes Physical Chemical Method SW-846, 3rd Edition, Updated IV December 1996
4. 1987 ASTM Annual Standards
5. Code of Federal Regulations, Title 40, Part 136, Appendix A, Revised July 1995
6. Methods for the Determination of Organic Compounds in Drinking Water, EPA-600/4-88/039, Revised July 1991, August 1995
7. NIOSH Manual of Anaytical Methods, 4th Edition, May 1996

GUARDIAN SYSTEMS, INC.

1108 Ashville Road
P.O. Box 190
Leeds, Alabama 35094

Telephone (205) 699-6647
Toll Free (866) 729-7211
Fax (205) 699-3882

City of Leeds
City of Leeds
1040 Park Drive
Leeds, AL 35094
Attention: Mr. Brad Watson

Report Date: 01/18/2019
Receive Date: 01/10/2019
Receive Time: 15:10

Control No : 1901-00156 Sample # 002
Sampler : DB
Sample ID: DC-SS-4

Sample Date: 01/10/2019
Sample Time: 9:43

Laboratory Certificate

PARAMETER	RESULTS	UNITS	ANALYST	DATE	TIME	METHOD	REF
pH	7.31	SU	DB	01/10/2019	9:43	150.1	(1)
Turbidity, Nephelometric	10.52	NTU	DB	01/10/2019	9:43	180.1	
Conductance, Sp. (field)	207	umhos	DB	01/10/2019	9:43	120.1	(1)
Nitrogen, Ammonia	<0.1	mg/L	TLB	01/16/2019	15:41	350.2	(1)
Color, APHA	58.	PCU	AH	01/11/2019	14:30	110.2	(1)
Hardness, Total	96.	mg/L	DRH	01/14/2019	12:00	200.7	
Calcium	30.6	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Magnesium	4.77	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Potassium	1.51	mg/L	DRH	01/14/2019	12:00	200.7	(1)

This Certificate is Continued On Next Page.

METHOD REFERENCES

1. Methods for Chemical Analysis of Water and Wastes. EPA-600/4-79-20, revised March 1983, August 1993 May 1994
2. Standard Methods for the Examination of Water and Waste Water, 18th, 19th, 20th, and 22nd Edition, 2012
3. Test Methods for Evaluating Solid Wastes Physical Chemical Method SW-846, 3rd Edition, Updated IV December 1996
4. 1987 ASTM Annual Standards
5. Code of Federal Regulations, Title 40, Part 136, Appendix A, Revised July 1995
6. Methods for the Determination of Organic Compounds in Drinking Water, EPA-600/4-88/039, Revised July 1991, August 1995
7. NIOSH Manual of Anaytical Methods, 4th Edition, May 1996

GUARDIAN SYSTEMS, INC.

1108 Ashville Road
P.O. Box 190
Leeds, Alabama 35094

Telephone (205) 699-6647
Toll Free (866) 729-7211
Fax (205) 699-3882

City of Leeds
City of Leeds
1040 Park Drive
Leeds, AL 35094
Attention: Mr. Brad Watson

Report Date: 01/18/2019
Receive Date: 01/10/2019
Receive Time: 15:10

Control No : 1901-00156 Sample # 003
Sampler : DB
Sample ID: LC-MP-3

Sample Date: 01/10/2019
Sample Time: 9:50

Laboratory Certificate

PARAMETER	RESULTS	UNITS	ANALYST	DATE	TIME	METHOD	REF
pH	7.45	SU	DB	01/10/2019	9:50	150.1	(1)
Turbidity, Nephelometric	3.40	NTU	DB	01/10/2019	9:50	180.1	
Conductance, Sp. (field)	407.	umhos	DB	01/10/2019	9:50	120.1	(1)
Nitrogen, Ammonia	<0.1	mg/L	TLB	01/16/2019	15:41	350.2	(1)
Color, APHA	20.	PCU	AH	01/11/2019	14:30	110.2	(1)
Hardness, Total	198.	mg/L	DRH	01/14/2019	12:00	200.7	
Calcium	69.0	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Magnesium	6.22	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Potassium	1.48	mg/L	DRH	01/14/2019	12:00	200.7	(1)

This Certificate is Continued On Next Page.

METHOD REFERENCES

1. Methods for Chemical Analysis of Water and Wastes. EPA-600/4-79-20, revised March 1983, August 1993 May 1994
2. Standard Methods for the Examination of Water and Waste Water, 18th, 19th, 20th, and 22nd Edition, 2012
3. Test Methods for Evaluating Solid Wastes Physical Chemical Method SW-846, 3rd Edition, Updated IV December 1996
4. 1987 ASTM Annual Standards
5. Code of Federal Regulations, Title 40, Part 136, Appendix A, Revised July 1995
6. Methods for the Determination of Organic Compounds in Drinking Water, EPA-600/4-88/039, Revised July 1991, August 1995
7. NIOSH Manual of Anaytical Methods, 4th Edition, May 1996

GUARDIAN SYSTEMS, INC.

1108 Ashville Road
P.O. Box 190
Leeds, Alabama 35094

Telephone (205) 699-6647
Toll Free (866) 729-7211
Fax (205) 699-3882

City of Leeds
City of Leeds
1040 Park Drive
Leeds, AL 35094
Attention: Mr. Brad Watson

Report Date: 01/18/2019
Receive Date: 01/10/2019
Receive Time: 15:10

Control No : 1901-00156 Sample # 004
Sampler : DB
Sample ID: LC-MP-2

Sample Date: 01/10/2019
Sample Time: 13:50

Laboratory Certificate

PARAMETER	RESULTS	UNITS	ANALYST	DATE	TIME	METHOD	REF
pH	7.25	SU	DB	01/10/2019	13:50	150.1	(1)
Turbidity, Nephelometric	4.40	NTU	DB	01/10/2019	13:50	180.1	
Conductance, Sp. (field)	397.	umhos	DB	01/10/2019	13:50	120.1	(1)
Nitrogen, Ammonia	<0.1	mg/L	TLB	01/16/2019	15:41	350.2	(1)
Color, APHA	19.	PCU	AH	01/11/2019	14:30	110.2	(1)
Hardness, Total	196.	mg/L	DRH	01/14/2019	12:00	200.7	
Calcium	68.1	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Magnesium	6.27	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Potassium	1.48	mg/L	DRH	01/14/2019	12:00	200.7	(1)

This Certificate is Continued On Next Page.

METHOD REFERENCES

1. Methods for Chemical Analysis of Water and Wastes. EPA-600/4-79-20, revised March 1983, August 1993 May 1994
2. Standard Methods for the Examination of Water and Waste Water, 18th, 19th, 20th, and 22nd Edition, 2012
3. Test Methods for Evaluating Solid Wastes Physical Chemical Method SW-846, 3rd Edition, Updated IV December 1996
4. 1987 ASTM Annual Standards
5. Code of Federal Regulations, Title 40, Part 136, Appendix A, Revised July 1995
6. Methods for the Determination of Organic Compounds in Drinking Water, EPA-600/4-88/039, Revised July 1991, August 1995
7. NIOSH Manual of Anaytical Methods, 4th Edition, May 1996

GUARDIAN SYSTEMS, INC.

1108 Ashville Road
P.O. Box 190
Leeds, Alabama 35094

Telephone (205) 699-6647
Toll Free (866) 729-7211
Fax (205) 699-3882

City of Leeds
City of Leeds
1040 Park Drive
Leeds, AL 35094
Attention: Mr. Brad Watson

Report Date: 01/18/2019
Receive Date: 01/10/2019
Receive Time: 15:10

Control No : 1901-00156 Sample # 005
Sampler : DB
Sample ID: LC-MP-1

Sample Date: 01/10/2019
Sample Time: 13:58

Laboratory Certificate

PARAMETER	RESULTS	UNITS	ANALYST	DATE	TIME	METHOD	REF
pH	7.62	SU	DB	01/10/2019	13:58	150.1	(1)
Turbidity, Nephelometric	2.47	NTU	DB	01/10/2019	13:58	180.1	
Conductance, Sp. (field)	386.	umhos	DB	01/10/2019	13:58	120.1	(1)
Nitrogen, Ammonia	<0.1	mg/L	TLB	01/16/2019	15:41	350.2	(1)
Color, APHA	17.	PCU	AH	01/11/2019	14:30	110.2	(1)
Hardness, Total	190.	mg/L	DRH	01/14/2019	12:00	200.7	
Calcium	66.4	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Magnesium	5.83	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Potassium	1.41	mg/L	DRH	01/14/2019	12:00	200.7	(1)

This Certificate is Continued On Next Page.

METHOD REFERENCES

1. Methods for Chemical Analysis of Water and Wastes. EPA-600/4-79-20, revised March 1983, August 1993 May 1994
2. Standard Methods for the Examination of Water and Waste Water, 18th, 19th, 20th, and 22nd Edition, 2012
3. Test Methods for Evaluating Solid Wastes Physical Chemical Method SW-846, 3rd Edition, Updated IV December 1996
4. 1987 ASTM Annual Standards
5. Code of Federal Regulations, Title 40, Part 136, Appendix A, Revised July 1995
6. Methods for the Determination of Organic Compounds in Drinking Water, EPA-600/4-88/039, Revised July 1991, August 1995
7. NIOSH Manual of Anaytical Methods, 4th Edition, May 1996

GUARDIAN SYSTEMS, INC.

1108 Ashville Road
P.O. Box 190
Leeds, Alabama 35094

Telephone (205) 699-6647
Toll Free (866) 729-7211
Fax (205) 699-3882

City of Leeds
City of Leeds
1040 Park Drive
Leeds, AL 35094
Attention: Mr. Brad Watson

Report Date: 01/18/2019
Receive Date: 01/10/2019
Receive Time: 15:10

Control No : 1901-00156 Sample # 006
Sampler : DB
Sample ID: LC-SS-6

Sample Date: 01/10/2019
Sample Time: 14:12

Laboratory Certificate

PARAMETER	RESULTS	UNITS	ANALYST	DATE	TIME	METHOD	REF
pH	7.43	SU	DB	01/10/2019	14:12	150.1	(1)
Turbidity, Nephelometric	2.41	NTU	DB	01/10/2019	14:12	180.1	
Conductance, Sp. (field)	369.	umhos	DB	01/10/2019	14:12	120.1	(1)
Nitrogen, Ammonia	<0.1	mg/L	TLB	01/16/2019	15:41	350.2	(1)
Color, APHA	18.	PCU	AH	01/11/2019	14:30	110.2	(1)
Hardness, Total	183.	mg/L	DRH	01/14/2019	12:00	200.7	
Calcium	64.2	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Magnesium	5.51	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Potassium	1.30	mg/L	DRH	01/14/2019	12:00	200.7	(1)

This Certificate is Continued On Next Page.

METHOD REFERENCES

1. Methods for Chemical Analysis of Water and Wastes. EPA-600/4-79-20, revised March 1983, August 1993 May 1994
2. Standard Methods for the Examination of Water and Waste Water, 18th, 19th, 20th, and 22nd Edition, 2012
3. Test Methods for Evaluating Solid Wastes Physical Chemical Method SW-846, 3rd Edition, Updated IV December 1996
4. 1987 ASTM Annual Standards
5. Code of Federal Regulations, Title 40, Part 136, Appendix A, Revised July 1995
6. Methods for the Determination of Organic Compounds in Drinking Water, EPA-600/4-88/039, Revised July 1991, August 1995
7. NIOSH Manual of Anaytical Methods, 4th Edition, May 1996

GUARDIAN SYSTEMS, INC.

1108 Ashville Road
P.O. Box 190
Leeds, Alabama 35094

Telephone (205) 699-6647
Toll Free (866) 729-7211
Fax (205) 699-3882

City of Leeds
City of Leeds
1040 Park Drive
Leeds, AL 35094
Attention: Mr. Brad Watson

Report Date: 01/18/2019
Receive Date: 01/10/2019
Receive Time: 15:10

Control No : 1901-00156 Sample # 007
Sampler : DB
Sample ID: LC-SS-7

Sample Date: 01/10/2019
Sample Time: 14:19

Laboratory Certificate

PARAMETER	RESULTS	UNITS	ANALYST	DATE	TIME	METHOD	REF
pH	7.77	SU	DB	01/10/2019	14:19	150.1	(1)
Turbidity, Nephelometric	2.55	NTU	DB	01/10/2019	14:19	180.1	
Conductance, Sp. (field)	360.	umhos	DB	01/10/2019	14:19	120.1	(1)
Nitrogen, Ammonia	<0.1	mg/L	TLB	01/16/2019	15:41	350.2	(1)
Color, APHA	26.	PCU	AH	01/11/2019	14:30	110.2	(1)
Hardness, Total	189.	mg/L	DRH	01/14/2019	12:00	200.7	
Calcium	67.0	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Magnesium	5.34	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Potassium	1.15	mg/L	DRH	01/14/2019	12:00	200.7	(1)

This Certificate is Continued On Next Page.

METHOD REFERENCES

1. Methods for Chemical Analysis of Water and Wastes. EPA-600/4-79-20, revised March 1983, August 1993 May 1994
2. Standard Methods for the Examination of Water and Waste Water, 18th, 19th, 20th, and 22nd Edition, 2012
3. Test Methods for Evaluating Solid Wastes Physical Chemical Method SW-846, 3rd Edition, Updated IV December 1996
4. 1987 ASTM Annual Standards
5. Code of Federal Regulations, Title 40, Part 136, Appendix A, Revised July 1995
6. Methods for the Determination of Organic Compounds in Drinking Water, EPA-600/4-88/039, Revised July 1991, August 1995
7. NIOSH Manual of Anaytical Methods, 4th Edition, May 1996

GUARDIAN SYSTEMS, INC.

1108 Ashville Road
P.O. Box 190
Leeds, Alabama 35094

Telephone (205) 699-6647
Toll Free (866) 729-7211
Fax (205) 699-3882

City of Leeds
City of Leeds
1040 Park Drive
Leeds, AL 35094
Attention: Mr. Brad Watson

Report Date: 01/18/2019
Receive Date: 01/10/2019
Receive Time: 15:10

Control No : 1901-00156 Sample # 008
Sampler : DB
Sample ID: LC-SS-9

Sample Date: 01/10/2019
Sample Time: 1:44

Laboratory Certificate

PARAMETER	RESULTS	UNITS	ANALYST	DATE	TIME	METHOD	REF
pH	7.73	SU	DB	01/10/2019	1:44	150.1	(1)
Turbidity, Nephelometric	3.57	NTU	DB	01/10/2019	1:44	180.1	
Conductance, Sp. (field)	352.	umhos	DB	01/10/2019	1:44	120.1	(1)
Color, APHA	24.	PCU	AH	01/11/2019	14:30	110.2	(1)
Nitrogen, Ammonia	<0.1	mg/L	TLB	01/16/2019	15:41	350.2	(1)
Calcium	58.9	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Magnesium	5.08	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Potassium	1.46	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Hardness, Total	168.	mg/L	DRH	01/14/2019	12:00	200.7	

This Certificate is Continued On Next Page.

METHOD REFERENCES

1. Methods for Chemical Analysis of Water and Wastes. EPA-600/4-79-20, revised March 1983, August 1993 May 1994
2. Standard Methods for the Examination of Water and Waste Water, 18th, 19th, 20th, and 22nd Edition, 2012
3. Test Methods for Evaluating Solid Wastes Physical Chemical Method SW-846, 3rd Edition, Updated IV December 1996
4. 1987 ASTM Annual Standards
5. Code of Federal Regulations, Title 40, Part 136, Appendix A, Revised July 1995
6. Methods for the Determination of Organic Compounds in Drinking Water, EPA-600/4-88/039, Revised July 1991, August 1995
7. NIOSH Manual of Anaytical Methods, 4th Edition, May 1996

GUARDIAN SYSTEMS, INC.

1108 Ashville Road
P.O. Box 190
Leeds, Alabama 35094

Telephone (205) 699-6647
Toll Free (866) 729-7211
Fax (205) 699-3882

City of Leeds
City of Leeds
1040 Park Drive
Leeds, AL 35094
Attention: Mr. Brad Watson

Report Date: 01/18/2019
Receive Date: 01/10/2019
Receive Time: 15:10

Control No : 1901-00156 Sample # 009
Sampler : DB
Sample ID: LC-SS-10

Sample Date: 01/10/2019
Sample Time: 14:52

Laboratory Certificate

PARAMETER	RESULTS	UNITS	ANALYST	DATE	TIME	METHOD	REF
pH	7.82	SU	DB	01/10/2019	14:52	150.1	(1)
Turbidity, Nephelometric	2.73	NTU	DB	01/10/2019	14:52	180.1	
Conductance, Sp. (field)	355.	umhos	DB	01/10/2019	14:52	120.1	(1)
Color, APHA	25.	PCU	AH	01/11/2019	14:30	110.2	(1)
Nitrogen, Ammonia	<0.1	mg/L	TLB	01/16/2019	15:41	350.2	(1)
Calcium	58.7	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Magnesium	5.17	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Potassium	1.49	mg/L	DRH	01/14/2019	12:00	200.7	(1)
Hardness, Total	168.	mg/L	DRH	01/14/2019	12:00	200.7	

Approved By:

MB

METHOD REFERENCES

1. Methods for Chemical Analysis of Water and Wastes. EPA-600/4-79-20, revised March 1983, August 1993 May 1994
2. Standard Methods for the Examination of Water and Waste Water, 18th, 19th, 20th, and 22nd Edition, 2012
3. Test Methods for Evaluating Solid Wastes Physical Chemical Method SW-846, 3rd Edition, Updated IV December 1996
4. 1987 ASTM Annual Standards
5. Code of Federal Regulations, Title 40, Part 136, Appendix A, Revised July 1995
6. Methods for the Determination of Organic Compounds in Drinking Water, EPA-600/4-88/039, Revised July 1991, August 1995
7. NIOSH Manual of Anaytical Methods, 4th Edition, May 1996

Guardian Systems, Inc.
1108 Ashville Road, P.O. Box 190
Leeds, Alabama 35094
(205) 699-6647
email: gsi@gsilab.com
(205) 699-3882 Fax
www.gsilab.com

Chain of Custody Record/ Analysis Report

Client:	Mr. Brad Watson	Phone:	
Company:	City of Leeds	Fax:	
Address:	1040 Park Drive	P.O.#:	
	Leeds, AL 35094	Project:	2019 1st Semester Stormwater- Dry

Sample ID	Sample Description	Sample Date	Sample Time	Sample: Comp.*	Sample: Grab	Bottle: Glass	Bottle: Plastic	Preservative: HCl	Preservative: HNO_3	Preservative: H_2SO_4	Preservative: NaOH	Preservative: Cool 4°C	Preservative: Other**	Analysis Requested					
	DC 553	1-10-19	9:47											Field pH:			7.11		
														Field Turbidity:			10.18		
														Field Conductivity:			~~254~~ µS		
					X		X					X		APHA Color			215		
					X		X			X		X		Ammonia					
					X		X					X		Hardness. Ca. Mg. K					
	DC 554	1-10-19	9:43											Field pH:			7.31		
														Field Turbidity:			10.52		
														Field Conductivity:			207 µS		
					X		X					X		APHA Color					
					X		X			X		X		Ammonia					
					X		X					X		Hardness. Ca. Mg. K					

Sampled by:	Dale Brakhage (PRINT)	(SIGN)	Relinquished by:	Date: 1-10-19	Time: 15:10
Received by:	Date:	Time:	Relinquished by:	Date:	Time:
Received by:	Date:	Time:	Relinquished by:	Date:	Time:
Received for Laboratory by: Lori Bryner	Date: 10 Jan 19	Time: 1510			
Was Shipped Container intact when received?	Yes ✓ No ___	Initials LB	Seals intact? Yes NA	No ___	
Were all samples properly preserved?	Yes ✓ No ___	Initials LB	Sample temp. 5.2 °C		
Comments:	Other = See Above for Sample Preservatives, Field Measurements				

Put an "X" in the appropriate column for sample type and sample preservative. Write in analysis requested.
* For composite samples include start and stop date and time in comments section **Write in preservative used in comments

Guardian Systems, Inc.
1108 Ashville Road, P.O. Box 190
Leeds, Alabama 35094
(205) 699-6647
email: gsi@gsilab.com

(205) 699-3882 Fax
www.gsilab.com

Chain of Custody Record/ Analysis Report

Client:	Mr. Brad Watson	Phone:	
Company:	City of Leeds	Fax:	
Address:	1040 Park Drive	P.O.#:	
	Leeds, AL 35094	Project:	2019 1st Semester Stormwater- Dry

Sample ID	Sample Description	Sample Date	Sample Time	Sample: Comp.*	Sample: Grab	Bottle: Glass	Bottle: Plastic	Preservative: HCl	Preservative: HNO_3	Preservative: H_2SO_4	Preservative: NaOH	Preservative: Cool 4°C	Preservative: Other**	Analysis Requested					
	LC-MP3	1-10-19	9:50											Field pH:			7.45		
														Field Turbidity:			3.40		
														Field Conductivity:			407 µS		
					X		X					X		APHA Color					
					X		X			X		X		Ammonia					
					X		X					X		Hardness. Ca. Mg. K					
	LC MP2	1-10-19	1:50											Field pH:			7.25		
														Field Turbidity:			4.40		
														Field Conductivity:			397 µS		
					X		X					X		APHA Color					
					X		X			X		X		Ammonia					
					X		X					X		Hardness. Ca. Mg. K					

Sampled by:	Dale Brakhage (PRINT)	(SIGN)		Relinquished by:		Date: 1-10-19	Time: 15:10
Received by:		Date:	Time:	Relinquished by:		Date:	Time:
Received by:		Date:	Time:	Relinquished by:		Date:	Time:
Received for Laboratory by:	Lori Bonner	Date: 10 Jan 19	Time: 1510				
Was Shipped Container intact when received?	Yes ✓	No ___	Initials LB	Seals intact? Yes NS	No ___		
Were all samples properly preserved?	Yes ✓	No ___	Initials LB	Sample temp. 5.2 °C			
Comments:	Other = See Above for Sample Preservatives, Field Measurements						

Put an "X" in the appropriate column for sample type and sample preservative. Write in analysis requested.
* For composite samples include start and stop date and time in comments section **Write in preservative used in comments

Guardian Systems, Inc.
1108 Ashville Road, P.O. Box 190
Leeds, Alabama 35094
(205) 699-6647
email: gsi@gsilab.com
(205) 699-3882 Fax
www.gsilab.com

Chain of Custody Record/ Analysis Report

Client:	Mr. Brad Watson	Phone:	
Company:	City of Leeds	Fax:	
Address:	1040 Park Drive	P.O.#:	
	Leeds, AL 35094	Project:	2019 1st Semester Stormwater- Dry

Sample ID	Sample Description	Sample Date	Sample Time	Sample: Comp.*	Sample: Grab	Bottle: Glass	Bottle: Plastic	Preservative: HCl	Preservative: HNO_3	Preservative: H_2SO_4	Preservative: NaOH	Preservative: Cool 4°C	Preservative: Other**	Analysis Requested		
	LC MP1	1-10-19	1:58											Field pH:	7.62	
														Field Turbidity:	2.47	
														Field Conductivity:	386 µS	
					X		X					X		APHA Color		
					X		X			X		X		Ammonia		
					X		X					X		Hardness. Ca. Mg. K		
	LC SS-6	1-10-19	2:12											Field pH:	7.43	
														Field Turbidity:	2.41	
														Field Conductivity:	369 µS	
					X		X					X		APHA Color		
					X		X			X		X		Ammonia		
					X		X					X		Hardness. Ca. Mg. K		

Sampled by: Dale Brakhage (PRINT) _____ (SIGN) Relinquished by: _____ Date: 1-10-19 Time: 15:10

Received by: _____ Date: _____ Time: _____ Relinquished by: _____ Date: _____ Time: _____

Received by: _____ Date: _____ Time: _____ Relinquished by: _____ Date: _____ Time: _____

Received for Laboratory by: _____ Date: 10 Jan 19 Time: 15:10

Was Shipped Container intact when received? Yes ✓ No ___ Initials LB Seals intact? Yes NA No ___

Were all samples properly preserved? Yes ✓ No ___ Initials LB Sample temp. 3.2 °C

Comments: Other = See Above for Sample Preservatives, Field Measurements

Put an "X" in the appropriate column for sample type and sample preservative. Write in analysis requested.
* For composite samples include start and stop date and time in comments section **Write in preservative used in comments

Guardian Systems, Inc.
1108 Ashville Road, P.O. Box 190
Leeds, Alabama 35094
(205) 699-6647
email: gsi@gsilab.com

(205) 699-3882 Fax
www.gsilab.com

Chain of Custody Record/ Analysis Report

Client:	Mr. Brad Watson	Phone:	
Company:	City of Leeds	Fax:	
Address:	1040 Park Drive	P.O.#:	
	Leeds, AL 35094	Project:	2019 1st Semester Stormwater- Dry

Sample ID	Sample Description	Sample Date	Sample Time	Sample: Comp.*	Sample: Grab	Bottle: Glass	Bottle: Plastic	Preservative: HCl	Preservative: HNO_3	Preservative: H_2SO_4	Preservative: NaOH	Preservative: Cool 4°C	Preservative: Other**	Analysis Requested	
	LC-357	1-10-19	2:19											Field pH:	7.77
														Field Turbidity:	2.55
														Field Conductivity:	360uS
					X		X					X		APHA Color	
					X		X			X		X		Ammonia	
					X		X					X		Hardness. Ca. Mg. K	
	LC SS 9	1-10-19	2140											Field pH:	7.73
														Field Turbidity:	3.57
														Field Conductivity:	352uS
					X		X					X		APHA Color	
					X		X			X		X		Ammonia	
					X		X					X		Hardness. Ca. Mg. K	

Sampled by: Dale Bradley PRINT Dale Bradley SIGN — Relinquished by: Dale Bradley — Date: 1-10-19 — Time: 15:10

Received by: — Date: — Time: — Relinquished by: — Date: — Time:

Received by: — Date: — Time: — Relinquished by: — Date: — Time:

Received for Laboratory by: Lori Brasher — Date: 10 Jan 19 — Time: 1510

Was Shipped Container intact when received? Yes / No ___ Initials LB — Seals intact? Yes NA No ___

Were all samples properly preserved? Yes / No ___ Initials LB — Sample temp. 5.2 °C

Comments: Other = See Above for Sample Preservatives, Field Measurements

Put an "X" in the appropriate column for sample type and sample preservative. Write in analysis requested.
* For composite samples include start and stop date and time in comments section **Write in preservative used in comments

Guardian Systems, Inc.
1108 Ashville Road, P.O. Box 190
Leeds, Alabama 35094
(205) 699-6647
email: gsi@gsilab.com
(205) 699-3882 Fax
www.gsilab.com

Chain of Custody Record/ Analysis Report

Client:	Mr. Brad Watson	Phone:	
Company:	City of Leeds	Fax:	
Address:	1040 Park Drive	P.O.#:	
	Leeds, AL 35094	Project:	2019 1st Semester Stormwater- Dry

Sample ID	Sample Description	Sample Date	Sample Time	Sample: Comp.*	Sample: Grab	Bottle: Glass	Bottle: Plastic	Preservative: HCl	Preservative: HNO_3	Preservative: H_2SO_4	Preservative: NaOH	Preservative: Cool 4°C	Preservative: Other**	Analysis Requested
	LC-5510	1-10-19	2:52											Field pH: 7.82
														Field Turbidity: 2.73
														Field Conductivity: 355 uS
					X		X					X		APHA Color
					X		X			X		X		Ammonia
					X		X					X		Hardness. Ca. Mg. K
														Field pH:
														Field Turbidity:
														Field Conductivity:
					X		X					X		APHA Color
					X		X			X		X		Ammonia
					X		X					X		Hardness. Ca. Mg. K

Sampled by:	Dale Brakhage (PRINT)	(SIGN)	Relinquished by:	Date: 1-10-19	Time: 15:10
Received by:	Date:	Time:	Relinquished by:	Date:	Time:
Received by:	Date:	Time:	Relinquished by:	Date:	Time:
Received for Laboratory by: Lori Bryner	Date: 10 Jan 19	Time: 1510			
Was Shipped Container intact when received?	Yes ✓ No ___	Initials LB	Seals intact? Yes NA No ___		
Were all samples properly preserved?	Yes ✓ No ___	Initials LB	Sample temp. 5.2 °C		
Comments:	Other = See Above for Sample Preservatives, Field Measurements				

Put an "X" in the appropriate column for sample type and sample preservative. Write in analysis requested.
* For composite samples include start and stop date and time in comments section **Write in preservative used in comments